



प्रतिभूति कागज कारखाना, नर्मदापुरम-461005 (म.प्र.)

(भारत प्रतिभूति मुद्रण तथा मुद्रा निर्माण निगम लिमिटेड की इकाई)

भारत सरकार के पूर्ण स्वामित्वाधीन

मिनीरल श्रेणी - 1 सीपीएसई एवं आई.एस.ओ. 9001 : 2015, 14001:2015, 45001:2018, 50001:2018 एवं आईईसी 17025:2017 प्रमाणित

SECURITY PAPER MILL, NARMADAPURAM - 461005 (M.P.)

(A Unit of Security Printing & Minting Corporation of India Limited)

Wholly Owned by Government of India

Miniratna Category - I CPSE & ISO 9001 : 2015, 14001:2015, 45001:2018, 50001:2018 & IEC 17025:2017 Certified

CIN : U22213DL2006GOI144763, GSTIN : 23AAJCS6111J3ZE



Tel. No. 07574-255259, Fax No. : 07574-255170, E-mail : gm.spm@spmcl.com, Website : <http://spmarmadapuram.spmcl.com>

शुद्धि पत्र - 1 / CORRIGENDUM-I

क्र. जेम/2025/बी/6391800

No. GEM/2024/B/6391800

दिनांक 04.08.2025

Dated 04.08.2025

सन्दर्भ : निविदा सूचना क्र. जेम/2025/बी/6391800, दिनांक 05.07.2025

Ref. : Tender Number: GEM/2024/B/6391800, Dated 05.07.2025

उपरोक्त संदर्भित निविदा सूचना दिनांक 05.07.2025 को हमारी वेबसाइट <https://spmarmadapuram.spmcl.com> पर प्रकाशित की गयी थी, जिसके संदर्भ में निम्नानुसार संशोधन करते हुए शुद्धि पत्र-1 जारी किया जा रहा है :

Against our above tender enquiry was published in the <https://spmarmadapuram.spmcl.com> on dated 05.07.2025. Further, to this **Corrigendum - I** as below is being issued to the tender enquiry with the following amendments:

Technical specification Sr No.	Existing Entry	Read As
1	The firm shall supply latest technology High Resolution Video Spectral Comparator System with DUAL Camera (top and bottom) for Examination of various security/currency papers and Forensic Documents.	The firm shall supply latest technology High Resolution Video Spectral Comparator System with Dual Camera (Top/primary & bottom/Secondary) for Examination of various security/currency papers and Forensic Documents.
7	Primary camera i.e. top camera should be of Colour, IR high resolution, high sensitivity of minimum 12MP with minimum resolution of 4000X3000 Pixel.	Primary / top camera should be of Colour, IR high resolution, high sensitivity of minimum 12MP with minimum resolution of 4000X3000 Pixel.
8	Secondary camera i.e. bottom camera should be of minimum 12MP with minimum resolution of 4000X3000 Pixel with enhanced field of view.	Bottom/Secondary Camera should be of minimum 12MP with minimum resolution of 4000X3000 Pixel with enhanced field of view.
32	Top and bottom camera should have IR, UV and white lights / filters.	Top/primary & bottom/Secondary Camera should have IR, UV and white lights/ filters.
58 (b)	Firm shall provide calibration micro scale for internal calibration purpose. It should be supplied with its traceability certificate issued by NABL laboratory.	Firm shall provide calibration micro scale for internal calibration purpose. It should be supplied with its traceability certificate issued by NABL laboratory/NIST traceable.
Bid End Date/Time	05.08.2025, 15:00:00	19.08.2025, 15:00:00
Bid opening Date/ Time	05.08.2025, 15:30:00	19.08.2025, 15:30:00

निविदा सूचना की अन्य सभी शर्तें यथावत रहेंगी ।

All other terms & condition of tender enquiry shall remain unchanged

(बिप्लब बसाक) / (Biplab Basak)

प्रबंधक (सामग्री)-(प्रमुख)/Manager (Material)-Material Head

हेतु मुख्य महाप्रबंधक/ For- Chief General Manager